CALIBRATION FOILS FOR COATING THICKNESS GAUGES

INSPECTION CERTIFICATE

- To calibrate or check the accuracy of coating thickness gauges
- Increase the measurement accuracy after foil calibration
- Uniformity: ±0.3μm (≤50μm), ±0.6% (>50μm)
- Dimension: 45×30mm (thickness≤2000μm), 50×30mm (thickness>2000μm)
- Supplied with manufacturer inspection certificate

Code	Thickness	Deviation*	Uniformity
9503-FL1	11.5μm	±5%	±0.3μm
9503-FL2	23μm	±5%	±0.3μm
9503-FL3	50.5μm	±5%	±0.6%
9503-FL4	102μm	±5%	±0.6%
9503-FL5	250μm	±5%	±0.6%
9503-FL6	495μm	±3%	±0.6%
9503-FL7	995μm	±2%	±0.6%
9503-FL8	1780μm	±5%	±0.6%
9503-FL9	3725μm	±4%	±0.6%
9503-FL10	7650μm	±2%	±0.6%
9503-FL11	780μm	±5%	±0.6%
9503-FL12	1400μm	±3%	±0.6%

* Deviation between the ordered thickness and the thickness marked on the calibration foils when supplied

9503-FL6